

Boston® Materials Recommendation

Application/Usage	Boston ES	Boston EO	Boston XO	Boston XO2
Daily wear	Excellent	Excellent	Excellent	Excellent
Planned replacement	Not recommended	Not recommended	Excellent	Excellent
Thin designs	Excellent	Excellent	Not recommended	Not recommended
Ultrathin designs	Excellent	Not recommended	Not recommended	Not recommended
Toric designs	Excellent (Hyperopic toric)	Excellent	Excellent (Hyperopic toric)	Excellent
Presbyopic designs	Excellent	Excellent	Excellent	Excellent
Corneal rehabilitation (post PMMA wear)	Excellent	Good	Good	Good
GPC rehabilitation	Excellent	Excellent	Excellent	Excellent
Keratoconus	Excellent	Excellent	Excellent	Excellent
Aphakia	Fair	Good	Excellent	Excellent
Post corneal surgery fitting	Excellent	Excellent	Excellent	Excellent
High oxygen demand corneas	Fair	Excellent	Excellent	Excellent
Dry eye	Excellent	Excellent	Excellent	Excellent
Ortho-K	Not recommended	Not recommended	Excellent	Excellent